

STAMMOPUR RD 5

Instrument cleaner

for ultrasound • concentrate • alkaline •

for dental laboratory and practice



Cleaning and pre-cleaning of workpieces and dental instruments

Concentrate for the removal of impurities on workpieces such as crowns, implants, dentures from manufacture, as well as medical devices and tools from machining in the dental laboratory.

Also for manual cleaning and pre-cleaning, both in the ultrasonic bath and in the immersion bath process, of dental, surgical, invasive and non-invasive dental instruments, accessories and components of medical devices in dental practices, dental clinics, maxillofacial surgery and orthodontics.

Very high material compatibility, suitable for parts made of stainless steel, titanium, glass, porcelain, ceramics and plastic. In case of doubt, it must be tested in advance.

Not suitable for use on aluminium and alkali-sensitive materials such as copper, brass, tin, zinc.

Application in the ultrasonic bath (dosage • treatment time)

3 % • 3 – 10 min

Application without ultrasound (dosage • treatment time)

5 % • max. 60 min;

depends on the type and stubbornness of the contamination

The application is diluted with water of at least drinking quality and without further additives. Ensure that parts are fully immersed. After treatment, rinse the parts thoroughly with water and dry. Change the working solution immediately if any contamination is visible. Otherwise, replace it at least once each day.

Prepare cold for fresh blood and protein residues (protein fixation); solution can be heated up to 60° C for stubborn contamination.

Observe the product information at www.dr-stamm.de.

Only for professional users.

Alkaline (pH 10.9 at 1 % in DI water), biodegradable.

Ingredients

5–15% anionic tensides, 5–15% non-ionic tensides, 5–15% phenols, <5% sodium hydroxide, complexing agents, alcohol.

Hazard statements

H315 Causes skin irritation.
H318 Causes serious eye damage.

Precautionary statements

P280 Wear protective gloves/protective clothing/eye protection/face protection.
P305+P351 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if
+P338 present and easy to do. Continue rinsing.

Danger



UFI-Code: 8M00-Q0KU-P00P-FPU2

Serious incidents involving the product must be reported to DR.H.STAMM GmbH Chemische Fabrik and to the competent authority.

Basis-UDI-DI: ++G014STRD507YB

This instruction for use is a translation. The binding edition is the German version.
S 3080/07e – 2023/01en



DR.H.STAMM GmbH Chemische Fabrik
Heinrichstraße 3-4 • 12207 Berlin (Germany)
Tel.: +49 30768 80 280 • Fax: +49 30 773 46 99
www.dr-stamm.de • info@dr-stamm.de

EC safety data sheets and product information as PDF
download at www.dr-stamm.de