

STAMMOPUR R

Instrument Cleaner

for Ultrasound • Concentrate



Cleaning and pre-cleaning of medical instruments and accessories

Concentrate for manual cleaning and pre-cleaning in instrument reprocessing, both in the ultrasonic bath and in immersion bath procedures for

general, surgical, invasive and non-invasive medical instruments and accessories and components of medical devices,

in the CSSD, in the hospital, in the doctor's practice, and in medical foot care.

Very high material compatibility, suitable for parts made of metal, titanium, glass, porcelain, ceramics and plastic. In case of doubt, this must be tested in advance.

The suitability recommendations of Karl Storz GmbH and Richard Wolf GmbH are available for endoscopes and their accessories.

Application in the ultrasonic bath (dosage • treatment time)

2 % • 3 – 10 min

Application without ultrasound (dosage • treatment time)

3 – 5 % • max. 12 h;

depending on the type and stubbornness of the contamination

The application is carried out diluted with water of at least drinking water quality and without further additives. Ensure that parts are fully immersed. After treatment, rinse the parts thoroughly with water and dry. Change the working solution immediately if contamination is visible, but at least once daily. Prepare cold for fresh blood and protein residues (protein fixation); solution can be heated up to 60 °C for stubborn contaminants.

Observe the product information at www.dr-stamm.de.

Only for professional users.

Can be used as contact liquid (2 %) in indirect sonication for basic cleaning with STAMMOPUR GR in an ultrasonic bath.

Mildly alkaline (pH 9.5 at 1 % in DI water), biodegradable.

Ingredients

5 – 15 % non-ionic tensides, 5 – 15 % anionic tensides, complexing agents, inhibitors, alcohol, scent.

Hazard statements

H318 Causes serious eye damage.

Precautionary statements

P280 Wear protective gloves/protective clothing/eye protection/face protection.

P305+P351 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if

+P338 present and easy to do. Continue rinsing.

Danger



UFI-Code: XH00-60WF-D006-TC80

Serious incidents involving the product must be reported to DR.H.STAMM GmbH Chemische Fabrik and to the competent authority.

Basis-UDI-DI: ++G014STR064S

This instruction for use is a translation. The binding edition is the German version.

S 3020/07e – 2023/01en



DR.H.STAMM GmbH Chemische Fabrik
Heinrichstraße 3-4 • 12207 Berlin (Germany)
Tel.: +49 30768 80 280 • Fax: +49 30 773 46 99
www.dr-stamm.de • info@dr-stamm.de

EC safety data sheets and product information as PDF
download at www.dr-stamm.de